



# Certificate of Traceability

## Manufacturer Certification

This certifies that **Boditech Med Inc.**, using **Immuno-assay on ichroma Reader** has participated in and successfully completed the NGSP certification for manufacturers and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: **Reference Material Institute for Clinical Chemistry Standards (ReCCS) ASRL#1**

The system evaluated was:

Instrument: <b>ichroma Reader</b>	Calibrator Lot: <b>929601</b>
Reagent Lot: <b>AAREL07</b>	Calibrator Assigned Values: <b>5.4%, 7.4%, 10.4%, 13.5%</b>

Date of Certification: July 1, 2021

Certification Expires: July 1, 2022

NGSP Steering Committee Chair

NGSP Network Coordinator

SRL director/ supervisor