



Certificate of Traceability

Manufacturer Certification


This certifies that *Boditech Med Inc.*, using *Immuno-assay on ichroma II* has participated in and successfully completed the NGSP certification for manufacturers and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: Reference Material Institute for Clinical Chemistry Standards (ReCCS) ASRL#1

The system evaluated was:

Instrument: ichroma II	Calibrator Lot: 929601
Reagent Lot: AAREL07	Calibrator Assigned Values: 5.4%, 7.4%, 10.4%, 13.5%

Date of Certification: July 1, 2021

Certification Expires: July 1, 2022



NGSP Steering Committee Chair



NGSP Network Coordinator



SRL director/ supervisor