



ichroma™ Myoglobin

INTENDED USE

ichroma™ Myoglobin is a fluorescence Immunoassay (FIA) for the quantitative determination of Myoglobin in human whole blood/serum/plasma. It is useful as an aid in management and monitoring of acute myocardial infarction (AMI).

For *in vitro* diagnostic use only.

INTRODUCTION

Myoglobin is an iron- and oxygen-binding protein found in both skeletal and myocardial muscles. It acts as a transport protein and is involved in diffusion of oxygen in the muscle tissue. Myoglobin is a single-chain globular protein of 154 amino acids. It is composed of a central iron-containing 'Heme' which is enclosed in a compact bundle-like or prism-like arrangement formed by the eight right-handed α -helices^{1,2}. Being a cytoplasmic protein having low molecular weight (of 17,699 daltons), myoglobin is released into the serum more rapidly as compared to other cardiac markers upon damage to the myocardial cells. Serum concentration of myoglobin increases above the normal range as early as 1 hour after acute myocardial infarction (AMI), attains peak level in approximately 4 to 8 hours after the onset and normalize rapidly afterwards. Thus myoglobin is better suited as a cardiac marker for early diagnosis of AMI. However, the elevated myoglobin is not specific to AMI owing to its large quantities in skeletal muscles as well. Despite its low clinical specificity and weak predictive value towards AMI, myoglobin is still a promising cardiac marker when other markers such as Creatin Kinase Isoenzyme-MB (CK-MB) and Cardiac Troponin-I (cTn-I) as well as other indicators like clinical signs and ECG are taken into account for diagnosis/confirmation of AMI³⁻⁸.

PRINCIPLE

The test uses a sandwich immunodetection method; the detector antibody in buffer binds to antigen in sample, forming antigen-antibody complexes, and migrates onto nitrocellulose matrix to be captured by the other immobilized-antibody on test strip.

The more antigen in sample forms the more antigen-antibody complex and leads to stronger intensity of fluorescence signal on detector antibody, which is processed by instrument for ichroma™ tests to show myoglobin concentration in sample.

COMPONENTS

ichroma™ Myoglobin consists of 'cartridges', 'Detection Buffer Tubes' and an 'ID chip'.

- The cartridge contains a test strip, the membrane which has anti human myoglobin at the test line, while streptavidin at the control line.
- Each cartridge is individually sealed in an aluminum foil pouch containing a desiccant. 25 sealed cartridges are packed in a box which also contains an ID chip.
- The detection buffer contains anti human myoglobin-fluorescence conjugate, biotin-BSA-fluorescence conjugate, bovine serum albumin (BSA) as a stabilizer and sodium azide as a preservative in phosphate buffered saline.
- The detection buffer is pre-dispensed in a tube. 25 detection buffer tubes are packaged in a box and further packed in a Styrofoam box with ice-pack for the shipment.

WARNINGS AND PRECAUTIONS

- For *in vitro* diagnostic use only.
- Carefully follow the instructions and procedures described in this 'Instruction for use'.
- Use only fresh samples and avoid direct sunlight.
- Lot numbers of all the test components (cartridge, ID chip and detection buffer) must match each other.
- Do not interchange the test components between different lots or use the test components after the expiration date, either of which might yield misleading of test result(s).
- Do not reuse. A detection buffer tube should be used for processing one sample only. So should a cartridge.
- The cartridge should remain sealed in its original pouch before use. Do not use the cartridge, if it is damaged or already opened.
- Frozen sample should be thawed only once. For shipping, samples must be packed in accordance with the regulations. Sample with severe hemolytic and hyperlipidemia cannot be used and should be recollected.
- Just before use, allow the cartridge, detection buffer and sample to be at room temperature for approximately 30 minutes.
- **ichroma™ Myoglobin** as well as the instrument for ichroma™ tests should be used away from vibration and/or magnetic field. During normal usage, it can be noted that instrument for ichroma™ tests may produce minor vibration.
- Used detection buffer tubes, pipette tips and cartridges should be handled carefully and discarded by an appropriate method in accordance with relevant local regulations.
- An exposure to larger quantities of sodium azide may cause certain health issues like convulsions, low blood pressure and heart rate, loss of consciousness, lung injury and respiratory failure.
- **ichroma™ Myoglobin** will provide accurate and reliable results subject to the following conditions.
 - Use **ichroma™ Myoglobin** should be used only in conjunction with instrument for ichroma™ tests.
 - Any anticoagulants other than EDTA, Heparin, Sodium citrate should be avoided.

STORAGE AND STABILITY

- The cartridge is stable for 20 months (while sealed in an aluminum foil pouch) if stored at 4-30 °C.
- The detection buffer pre-dispensed in a tube is stable for 20 months if stored at 2-8 °C.
- After the cartridge pouch is opened, the test should be performed immediately.

LIMITATION OF THE TEST SYSTEM

- The test may yield false positive result(s) due to the cross-reactions and/or non-specific adhesion of certain sample components to the capture/detector antibodies.
- The test may yield false negative result. The non-responsiveness of the antigen to the antibodies is most common where the epitope is masked by some unknown components, so as not to be detected or captured by the antibodies. The instability or degradation of the antigen with time and/or temperature may cause the false negative as it makes antigen unrecognizable by the antibodies.
- Other factors may interfere with the test and cause erroneous results, such as technical/procedural errors, degradation of the test components/reagents or presence of interfering substances in the test samples.
- Any clinical diagnosis based on the test result must be supported by a comprehensive judgment of the concerned physician including clinical symptoms and other relevant test results.

MATERIALS SUPPLIED

REF CFPC-37

Components of **ichroma™ Myoglobin**

- Cartridge Box:
 - Cartridges 25
 - ID Chip 1
 - Instruction For Use 1
- Box containing Detection Buffer Tubes
 - Detection Buffer Tubes 25

MATERIALS REQUIRED BUT SUPPLIED ON DEMAND

Following items can be purchased separately from **ichroma™ Myoglobin**. Please contact our sales division for more information.

- Instrument for **ichroma™** tests
 - **ichroma™ Reader** **REF** FR203
 - **ichroma™ II** **REF** FPRR021
 - **ichroma™ D** **REF** 13303
- **ichroma™ Printer** **REF** FPRR007
- **Boditech Cardiac Control** **REF** CFPO-98

SAMPLE COLLECTION AND PROCESSING

The sample type for **ichroma™ Myoglobin** is human whole blood/serum/plasma.

- It is recommended to test the sample within 24 hours after collection.
- The serum or plasma should be separated from the clot by centrifugation within 3 hours after the collection of whole blood.
- Samples may be stored for up to a week at 2-8 °C prior to being tested. If testing will be delayed more than a week, samples should be frozen at -20 °C.
- Samples stored frozen at -20 °C for 3 months showed no performance difference.
- However, the whole blood sample should not be kept in a freezer in any case.
- Once the sample was frozen, it should be used one time only for test, because repeated freezing and thawing can result in the change of test values.

TEST SETUP

- Check the contents of **ichroma™ Myoglobin**: Sealed Cartridge, Detection Buffer Tube and ID Chip.
- Ensure that the lot number of the cartridge matches that of the ID chip as well as the detection buffer.
- Keep the sealed cartridge (if stored in refrigerator) and the detection buffer tube at room temperature for at least 30 minutes just prior to the test. Place the cartridge on a clean, dust-free and flat surface.
- Turn on the instrument for **ichroma™** tests.
- Insert the ID Chip into the ID chip port of the instrument for **ichroma™** tests.
- Press the 'Select' button on the instrument for **ichroma™** tests. (Please refer to the 'Instrument for **ichroma™** tests Operation Manual' for complete information and operating instructions.)

TEST PROCEDURE

- 1) Transfer 10 µL (Human whole blood/serum/plasma/control) of sample using a transfer pipette to a tube containing the detection buffer.
- 2) Close the lid of the detection buffer tube and mix the sample thoroughly by shaking it about 10 times.
- 3) Pipette out 75 µL of a sample mixture and load it into the sample well on the cartridge.
- 4) Leave the sample-loaded cartridge at room temperature for 12 minutes.

⚠ Scan the sample-loaded cartridge immediately when the incubation time is over. If not, it will cause inexact test result.

- 5) To scan the sample-loaded cartridge, insert it into the cartridge holder of the instrument for **ichroma™** tests. Ensure proper orientation of the cartridge before pushing it all the way inside the cartridge holder. An arrow has been marked on the cartridge especially for this purpose.
- 6) Press 'Select' button on the instrument for **ichroma™** tests to start the scanning process.
- 7) Instrument for **ichroma™** tests will start scanning the sample-loaded cartridge immediately.
- 8) Read the test result on the display screen of the instrument for **ichroma™** tests.

INTERPRETATION OF TEST RESULT

- Instrument for **ichroma™** tests calculates the test result automatically and displays myoglobin concentration of the test sample in terms of ng/mL.
- The cut-off (reference value): 70 ng/mL
- Working range : 5-500 ng/mL

QUALITY CONTROL

- Quality control tests are a part of the good testing practice to confirm the expected results and validity of the assay and should be performed at regular intervals.
- The control tests should be performed immediately after opening a new test lot to ensure the test performance is not altered.
- Quality control tests should also be performed whenever there is any question concerning the validity of the test results.
- Control materials are not provided with **ichroma™ Myoglobin**. For more information regarding obtaining the control materials, contact **Boditech Med Inc.'s Sales Division for assistance**. (Please refer to the instruction for use of control material.)

PERFORMANCE CHARACTERISTICS

- **Analytical Sensitivity**

Limit of Blank (LoB)	1.23 ng/mL
Limit of Detection (LoD)	1.86 ng/mL
Limit of Quantification (LoQ)	5.0 ng/mL
- **Analytical Specificity**

- Cross-reactivity
There was no significant cross-reactivity from these materials with the **ichroma™ Myoglobin** test measurements.

Material	Conc. Of Control [ng/mL]					
	10	55	180	10	55	180
	CV(%)			Recovery (%)		
Troponin complex	4.79	5.67	5.16	102	102	97
CK-MB	6.05	6.60	4.74	98	100	99
D-dimer	4.91	5.20	4.03	102	102	97
NT-proBNP	6.26	6.46	4.79	101	100	99

- Interference
There was no significant interference from these materials with the **ichroma™ Myoglobin** test measurements.

Material	Conc. Of Control [ng/mL]					
	10	55	180	10	55	180
	CV(%)			Recovery (%)		
Bilirubin	3.89	5.57	5.79	97.38	100.75	99.22
Cholesterol	6.48	6.25	5.04	100.12	101.35	99.50
D-glucose	4.64	5.29	5.46	102.69	98.51	102.28
Hemoglobin	6.72	5.80	7.06	98.60	100.47	99.35
L-Ascorbic acid	6.55	6.36	6.40	99.88	100.41	98.75
Triglyceride mixture	5.26	6.39	7.05	98.19	101.56	99.13

■ Precision

- Between lot
One person tested three different lots of **ichroma™ Myoglobin**, ten times at each concentration of the control standard.
- Between person
Three different persons tested **ichroma™ Myoglobin**, ten times at each concentration of the control standard.
- Between day
One person tested **ichroma™ Myoglobin** during five days; ten times at each concentration of the control standard.
- Between site
Three different person tested **ichroma™ Myoglobin** at three different sites; ten times at each concentration of the control standard.


conc. [ng/mL]	Between-lot		Between-person		Between-day		Between-site	
	AVG	CV(%)	AVG	CV(%)	AVG	CV(%)	AVG	CV(%)
10	9.92	4.8	9.89	5.8	9.93	6.5	9.91	6.1
55	53.92	5.5	54.71	5.4	54.62	6.1	55.78	5.2
180	177.11	5.3	183.17	5.5	182.01	6.0	180.16	6.2

■ Accuracy

The accuracy was confirmed by 3 different lots testing ten times each different concentrations.

conc. [ng/mL]	Lot 1	Lot 2	Lot 3	AVG	Recovery (%)
10.00	10.05	10.03	9.72	9.93	99.3
55.00	56.76	52.76	55.19	54.91	99.8
180.00	178.82	175.04	175.28	176.38	98.0

- **Comparability:** Myoglobin concentrations of 104 clinical samples were independently with **ichroma™ Myoglobin** and Mini VIDAS® (BioMérieux Inc. France) as per prescribed test procedures. Test results were compared and their comparability was investigated with linear regression and coefficient of correlation (R). Linear regression and coefficient of correlation between the two tests were $Y=0.9893X+0.4756$ and $R=0.9955$ respectively.



REFERENCES

1. Kent Lewandrowski, Ahchean Chen and James Januzzi, Cardiac markers for myocardial infarction, Am J Clin Pathol 2002;118 (Suppl 1):S93-S99.
2. Cox, MM, Nelson, DL. Lehninger: Principles of Biochemistry, 3rd edition. W.H. Freeman and Company, New York, 2000, 206.
3. Ordway GA, Garry DJ. Myoglobin: An essential hemoprotein in striated muscle. J Exp Biol. 2004;207(Pt 20):3441-6.
4. Lewandrowski K, Chen A, Januzzi J. Cardiac markers for myocardial infarction. A brief review. Am J Clin Pathol. 2002;118:S93-9.
5. Vaidya HC. Myoglobin: an early biochemical marker for the diagnosis of acute myocardial infarction. J Clin Immunoassay. 1994;17:35-39.

6. Gibler WB, Gibler CD, Weinshenker C, et al. Myoglobin as an early indicator of acute myocardial infarction. Ann Emerg Med. 1987;16:851-856.
7. Adams JE, Abendschein DR, Jaffe AS. Biochemical markers of myocardial injury: is MB creatine kinase the choice for the 1990s? Circulation. 1993;88:750-763.
8. Mair J, Morandell D, Genser N, et al. Equivalent early sensitivities of myoglobin, creatine kinase-MB mass, creatine kinase isoforms ratios, and cardiac troponins I and T for acute myocardial infarction. Clin Chem. 1995;41:1266-1272.
9. Mercer DW. Role of cardiac markers in evaluation of suspected myocardial infarction. Postgrad Med. 1997;102:113-122.

Note: Please refer to the table below to identify various symbols

	Sufficient for <n> tests
	Read instruction for use
	Use by Date
	Batch code
	Catalog number
	Caution
	Manufacturer
	Authorized representative of the European Community
	In vitro diagnostic medical device
	Temperature limit
	Do not reuse
	This product fulfills the requirements of the Directive 98/79/EC on in vitro diagnostic medical devices

For technical assistance; please contact:

Boditech Med Inc.'s Technical Services

Tel: +82 33 243-1400

E-mail: sales@boditech.co.kr



Boditech Med Incorporated

43, Geodudanji 1-gil, Dongnae-myeon,
Chuncheon-si, Gang-won-do, 24398
Republic of Korea

Tel: +(82) -33-243-1400

Fax: +(82) -33-243-9373

www.boditech.co.kr



Obelis s.a

Bd. Général Wahis 53,
1030 Brussels, BELGIUM

Tel: +(32) -2-732-59-54

Fax: +(32) -2-732-60-03

E-Mail: mail@obelis.net

